

ANALYTICAL CERTIFICATE

Page 1/10

| | |
|--|----------------------------------|
| Sample name | SS-31 |
| Batch No. | 2024252 |
| Sample No. | 01 |
| Sequence | DArg-Dmt-Lys-Phe-NH ₂ |
| Manufacturing date | NA |
| Submitter of analytical request | Adria Peptides d.o.o. |

1. Peptide content by HPLC/CLND:

1.1 HPLC Instrument:

Pump: Agilent 1200 Series, Quat Pump G1311A
Sampler: Agilent 1260 Series, Hip ALS G1367E
Degasser: Agilent 1200 Series, Degasser G1379B
Detectors: Agilent 1200 Series, VWD G1314B
Nitrogen detector Antek 8060

1.2 HPLC conditions:

Eluents: A – MilliQ water
B – isopropanol
D – 1% TFA in MilliQ water
Flow rate: 1 mL/min
Gradient:

| Time | A (%) | B (%) | D (%) |
|------|-------|-------|-------|
| 0 | 90 | 0 | 10 |
| 1 | 90 | 0 | 10 |
| 9 | 10 | 80 | 10 |
| 10 | 10 | 80 | 10 |
| 11 | 90 | 0 | 10 |
| 15 | 90 | 0 | 10 |

Column: ARION 5 μ C4-BIO 300 A, 4.6 x 100 mm
Serial No 221258


1.3 Sample preparation:

The whole amount of SS-31 (50 mg) was dissolved in 10 mL of MilliQ water.
Injection: 2.0 μ L

ANALYTICAL CERTIFICATE

Page 2/10

1.4 Chromatograms and calibration curve:




ANALYTICAL CERTIFICATE

Page 3/10

Chromatogram and Results

| Injection Details | | |
|----------------------|-------------------------|-------------------------|
| Injection Name: | SS-31 (PO-2024252) frc2 | Run Time (min): 16.00 |
| Vial Number: | P1:E1 | Injection Volume: 2.00 |
| Injection Type: | Unknown | Channel: CLND_Analog |
| Calibration Level: | | Wavelength: |
| Instrument Method: | AspartV10 long 9min | Bandwidth: n.a. |
| Processing Method: | Aspart V20 | Dilution Factor: 1.0000 |
| Injection Date/Time: | 23.5.25 12:13 | Sample Weight: 1.0000 |

Chromatogram

Integration Results

| No. | Peak Name | Retention Time min | Area mV*min | Height mV | Amount ug N | Concentration mg N/mL | Evaluation |
|---------------|------------|-----------------------|----------------|----------------|----------------|--------------------------|------------|
| 1 | Compound Y | 5.415 | 68.040 | 439.948 | 1.9575 | 0.9788 | valid |
| n.a. | Compound X | n.a. | n.a. | n.a. | n.a. | n.a. | n.a. |
| Total: | | | 68.040 | 439.948 | 1.96 | 0.98 | |


ANALYTICAL CERTIFICATE

Page 4/10

Calibration

| Calibration Details | | n.a. | |
|---------------------------------------|-----------------|-------------|------|
| Calibration Type | Lin, WithOffset | Offset (C0) | n.a. |
| Evaluation Type | Area | Slope (C1) | n.a. |
| Number of Calibration Points | n.a. | Curve (C2) | n.a. |
| Number of disabled Calibration Points | n.a. | R-Square | n.a. |

Calibration Plot



Calibration Results

| Calibration Results | | n.a. | | | | | |
|---------------------|----------------|-------------------|---------------------------|---------------------------|---------------------------|-------------------------------------|---------------------------------|
| No. | Injection Name | Calibration Level | X Value | Y Value | Y Value | Area | Height |
| | | | CLND_Analog Compound X | CLND_Analog Compound X | CLND_Analog Compound X | mV*min CLND_Analog Compound X | mV CLND_Analog Compound X |
| 2 | Aspart5 | 1 | 2.7408 | 95.1396 | 95.1396 | 95.140 | 555.193 |
| 3 | Aspart4 | 1 | 2.1926 | 76.7674 | 76.7674 | 76.767 | 441.850 |
| 4 | Aspart3 | 1 | 1.6445 | 57.0328 | 57.0328 | 57.033 | 329.763 |
| 5 | Aspart2 | 1 | 1.0963 | 38.0617 | 38.0617 | 38.062 | 215.626 |
| 6 | Aspart1 | 1 | 0.5482 | 18.2048 | 18.2048 | 18.205 | 103.476 |

ANALYTICAL CERTIFICATE

Page 5/10

1.4 Results:

| | | |
|--|--|--|
| NNC: SS-31 (PO-2024252) | | Salt: 0 |
| MW <i>(calculated)</i> g/mol | N content <i>(calculated)</i> % | N conc. <i>(measured)</i> mg × N/ml |
| 639,8 | 19,7 | 0,9759 |
| Theoretical Volume ml | | Lyophilizate amount mg |
| 10,00 | | 50,00 |
| Peptide concentration mg/ml nmol/ml | | Quantified amount mg nmol |
| 4,95 7742 | | 49,5 77 423 |
| Peptide content assay % | | |
| 99,1 | | |

Summary table:

| Peptide | Aliquoting (mg) | Total weight of sample (mg) | Content of the peptide by CLND (mg) | Content of the peptide in the sample (%) | Content of the peptide against the amount on label. |
|---------|-----------------|-----------------------------|-------------------------------------|--|---|
| SS-31 | 50 | -- | 49,54 | -- | 99,1 % |

ANALYTICAL CERTIFICATE

Page 6/10

2. Purity assessment by UPLC:

2.1 HPLC Instrument:

LC-System Waters Acquity UPLC
Detectors: UV or DAD at 214 nm

2.2 HPLC conditions:

Eluents: A – MilliQ water + 0.05% TFA
 B – acetonitrile + 0.05% TFA
Flow rate: 0.40 mL/min
Gradient: from 5% B to 60% B in 16 min, according to chromatogram results
Column: Waters Acquity BEH, C-18, 1.7 μ m, 2.1mm x 150mm

2.3 Sample preparation:

An aliquote of SS-31 (1 mg) was dissolved in 1 mL of water.
Injection: 1.0 μ L

ANALYTICAL CERTIFICATE

Page 7/10

2.4 Chromatogram of SS-31 (PO-2024252)

Sample information

UPLC5

Channel Description PDA Ch1 214nm@4.8nm

Vial : 1:A,3 Vol. : 1.00 ul


Sample: **SS-31 (PO-2024252)**

Date Acquired 5/26/2025 1:04:03 PM CEST

Date Processed 5/26/2025 2:53:36 PM CEST

Acq Method Set :

Gr5_60_4mi_40C_0_45ml_K2_met_s



| | RT | Area | Height (μV) | % Area |
|---|-------|---------|-------------|--------|
| 1 | 1.929 | 2599351 | 1477324 | 99.51 |

A: 0.05% TFA in water

B: 0.05% TFA in acetonitrile

Gradient :

0.0 - 0.5min 5 - 5 % B

0.5 - 4 min 5 - 60 % B

4.0 - 4.5 min 60 - 100 % B

4.5 - 5.0min 100 % B

5.0 - 5.5min 100 - 5 % B

6min 5 % B

0.45ml/min

Acquity UPLC BEHC18, 1.7μm, 2.1 x 50 mm column
column own temp. = 40 °C

2.5 Result of purity assessment

The overall purity is 99.51 % at 214 nm.

ANALYTICAL CERTIFICATE

Page 8/10

3. Peptide identity by UPLC/MS:

3.1 HPLC Instrument:

LC-System Waters Acquity UPLC
Detectors: UV or DAD at 214 nm

3.2 HPLC conditions:

Eluents: A – MilliQ water + 0.05% TFA
B – acetonitrile + 0.05% TFA
Flow rate: 0.40 mL/min
Gradient: from 5% B to 60% B in 4 min, according to chromatogram results
Column: Waters Acquity BEH, C-18, 1.7 μ m, 2.1mm x 50mm
Part No 186002353

3.3 MS Detector:

Detector Waters (Micromass) ZQ 2000
Ionisation method: ES+
Scanning range: 200 – 2000 amu
Capillary voltage: 3.0 kV
Cone Voltage: 20 V
Scantime: 0.9 s
Interscan delay: 0.1 s
Detection method: quadrupole

3.4 Sample preparation:

An aliquote of SS-31 (1 mg) was dissolved in 1 mL of water.
Injection: 0.8 μ L

ANALYTICAL CERTIFICATE

Page 9/10

3.5 Mass Spectra of SS-31 (PO-2024252)

Sample information

UPLC1 ZQ

Sample: **SS-31 (PO-2024252)**

Channel Description ACQUITY TUV ChA 214nm


Date Acquired 5/26/2025 1:10:13 PM CEST

Vial : 1:A,2 Vol. : 0.80 ul

Date Processed 5/26/2025 2:55:10 PM CEST

Acq Method Set :

Gr5_60_MS_4min_0_45ml_K2_me_s



MS Result Table

| Name | RT | Base Peak (m/z) |
|------|-------|-----------------|
| 1 | 1.793 | 320.82 |

UPLC conditions:

A: 0.05% TFA in water

B: 0.05% TFA in acetonitrile

Gradient :

0.0 - 0.5min 5 - 5 % B

0.5 - 4 min 5 - 60 % B

0.45ml/min

Acquity UPLC BEHC18, 1.7um, 2.1 x 50 mm column

column temp. = 40 °C

Theoretical values of m/z:

| Peptide MW | [M+1H] ²⁺ | [M+2H] ³⁺ | [M+3H] ⁴⁺ | [M+4H] ⁵⁺ | [M+5H] ⁶⁺ | [M+6H] ⁷⁺ |
|------------|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|
| 639,8 | 640,8 | 320,9 | | | | |
| Found | 640,5 | 320,8 | | | | |

ANALYTICAL CERTIFICATE

Page 10/10

4. Endotoxin test:

4.1 Description:

Test tubes: Gel Clot Lyophilized Amebocyte Lysate Single Test in Vial
 Manufacturer: Xiamen Bioendo Technology Co., Ltd.
 Lot: 24061152
 Content: 0.2 ml endotoxin-specific Amebocyte Lysate which includes beta-glucan inhibitor in the formulation
 Sensitivity of test: 0.5 EU/mL
 Sample sensitivity level: 5 EU/mg

4.2 Sample preparation and test:

A sample peptide is dissolved in endotoxin-free water to form a concentration of 0.5 mg/mL. 200 µL of this solution is then transferred to the Amebocyte Lysate Single Test tube and incubated at 37 °C for 60 min. Immediately after incubation the test tube is slowly turned upside down.

A solid gel clot which doesn't come down immediately indicates **positive** result (meaning that endotoxins are above the current sensitivity level).

An absence of solid gel clot so the solution freely flows down from the bottom of test tube indicates **negative** result (meaning that endotoxin are below the current sensitivity level).

4.3 Result:

NEGATIVE (-)

CONCLUSION:


The sample SS-31 (Batch No. 2024252) was analyzed for peptide content, acetic acid content, UV purity, identity by MS and endotoxins.

Peptide content is 99.08 % (49.54 mg in 50 mg)

Purity is 99.51 % (UPLC at 214 nm).

MS identity complies with theoretical calculation of m/z values.

Endotoxin test (sensitivity level 5 EU/mg) - NEGATIVE.

| | |
|----------------------------|---|
| | |
| ANALYSIS COMPLETED: | Date: 26.05.2025 |
| Issued by QC: | Date: 27.05.2025 Signature:  |

Certificate Of Analysis



Client:**Particle Peptides**

Viktor Mužila
Kolonáda 4490/18
98401 Lučenec
Slovakia
muzila@particlepeptides.com
+421 917 381 743

Laboratory:

Liquilabs s.r.o.
Ohradní 24B
14000 Michle
Czechia
www.liquilabs.cz


Sample Identification

| | | | | | |
|--------------------|-------------|---------------------|---------|-----------------------|------------------|
| Sample Name | SS-31 50 mg | Batch Number | 2025252 | Date Published | 2025-07-29 12:36 |
|--------------------|-------------|---------------------|---------|-----------------------|------------------|

Results for Lyo-0065

| Bioburden | Result | Unit | Uncertainty | Reporting Limit | |
|--|---------------|-------------|--------------------|------------------------|---|
| Total Aerobic Microbial Count USP <61> Plate Count Method | Not detected | CFU/g | | >= 1000 | △ |
| Total Yeast and Mold Count USP <61> Plate Count Method | Not detected | CFU/g | | >= 100 | △ |

Attachments for Lyo-0065

| | | |
|---|---------------------------------|-----------------------------|
|  | Method Specification | |
| Determination of bioburden of lyophilized samples | | |
| <i>Document number</i> MIC_001_2025 | <i>Superseded document</i> - | <i>Number of pages</i> 2 |

1. Instrumentation and chemicals

1.1. Instruments used

- Sterile Syringe 2mL Luer
- Sterile needles
- Ready made PCA Plate ROTI Aquatest
- Ready made Sab4 Plate ROTI Aquatest

1.2. Chemicals

Sterile physiological solution (0.9% NaCl)

2. Sample preparation and inoculation

2.1 Sample preparation

1. Fresh sterile needle and syringe was used for measuring exactly 2 mL of sterile physiological solution.
2. Needle was changed and by new needle rubber top of peptide container was penetrated and 2 mL of sterile physiological solution was dispensed.
3. Content of container was completely dissolved and left for 5 minutes to settle potentially created bubbles.
4. This procedure is repeated for two vials.

2.2 Total Aerobic microbial count inoculation and cultivation

1. By sterile needle 1 mL of solution was filled into the sterile syringe.
2. Needle was placed above the flame for few seconds to sterilize.
3. Consequently 1 mL of solution was poured into the ready to use sterile petri dish filled with PCA agar and petri dish was closed.
4. Proces was repeated for two petri dishes.
5. With sterile needle, 1 mL of sterile physiological solution was filled into the sterile needle and was inoculated onto one sterile petri dish filled with PCA agar as negative control sample.
6. Samples and negative control sample were placed in incubator at temperature 37°C for 120h.

1

Attachment for Lyo-0065
Filename: Bioburden-images-0.jpg

2.3 Total Yeast and Mold count inoculation and cultivation

1. By sterile needle 1 mL of solution was filled into the sterile syringe.
2. Needle was placed above the flame for few seconds to sterilize.
3. Consequently 1 mL of solution was poured into the ready to use sterile petri dish filled with Sab4 agar and petri dish was closed.
4. Proces was repeated for two petri dishes.
5. With sterile needle, 1 mL of sterile physiological solution was filled into the sterile needle and was inoculated onto one sterile petri dish filled with Sab4 agar as negative control sample.
6. Samples and negative control sample were placed in incubator at temperature 25°C for 72h.

3. Evaluation of results

After incubation time, colonies are counted as cfu (colonies forming units) and result per 1g of sample is determined as:

$$CFU_{avg} = \frac{\sum CFU_n}{n}$$

CFU_{avg} = average CFU counted from n inoculations

CFU_n = CFU counted per inoculation

n = number of inoculations

$$CFU \text{ per gram} = \frac{CFU_{avg}}{m_s} * DF$$

CFU_{avg} = Average CFU counted from n inoculations

m_s = mass of sample (mg)

DF = Dilution factor

If negative control sample is evaluated as positive, process have to be repeated due to possible contamination in the process of inoculation or incubation.

2

Attachment for Lyo-0065
Filename: Bioburden-images-1.jpg

Responsibles



Mr. Ján Galbavý
Founder/Manager

Analysis results relate only to the samples tested.

This document shall not be reproduced except in full, without the written approval of Liquilabs s.r.o.

Analytical report AR-25-KT-018095-02



| | |
|--|---|
| Testing laboratory: Eurofins Environment Testing Slovakia s.r.o. Robotnícka 820/36, 039 01 Turčianske Teplice IČO: 53 248 376 Place of work: Accredited testing laboratory Turčianske Teplice Robotnícka 820/36, 039 01 Turčianske Teplice tel: 043/490 1562 Registration EnviroSK@etcee.eurofins.com, www.eurofins.sk | Customer: PARTICLE s.r.o. Kolonáda 4490/18 984 01 Lučenec SLOVAKIA |
|--|---|

Date of Sample Receipt: 20.05.2025 Date of Testing: 20.05.2025 - 22.05.2025

Issue date: 22.05.2025

Information about Sampling:

Sampler: customer
Sample information: 104-2025-00019379
 # Sample description: SS-31 (PO-2024252)
 Material: Peptidy

Physical and chemical tests

| Parameter | Unit | Allowed Value | Measured Value | Uncertainty of Method measurement* | Method | Testing method | E | SL | TT |
|--------------|-------|---------------|----------------|------------------------------------|--------|----------------|---|----|----|
| Arsenic (As) | mg/kg | max, 1,5 | 0,039 | 25% | ICP-MS | LS-PP-CH-85 | S | TR | A |
| Cadmium (Cd) | mg/kg | max, 0,2 | <0,10 | - | ICP-MS | LS-PP-CH-85 | S | TR | A |
| Lead (Pb) | mg/kg | max, 0,5 | <0,30 | - | ICP-MS | LS-PP-CH-85 | S | TR | A |
| Mercury (Hg) | mg/kg | max, 0,3 | 0,018 | 25% | ICP-MS | LS-PP-CH-85 | S | TR | A |

Notes:

| | |
|---|---|
| E - evaluation | TT - type of test |
| S - satisfied | A - accredited test executed at the own test laboratory |
| NS - not satisfied | N - non accredited test executed at the own test laboratory |
| (A) - accredited sampling | SA - accredited test executed under the subcontract |
| (SA) - accredited sampling executed under the subcontract | SN - unaccredited test executed under the subcontract |
| ŠPP - Standard operation procedure | (TM) - testing outside the laboratory at the customer |
| ND - not detected by given method | |
| LOQ, LQ – limit of quantification | |
| CFU - Colony forming unit | |
| NM - necessary quantity | |
| m - the highest allowed value at the case of one sample | |
| M, c - "M" highest allowed value for the number "c" at the case of 5 sample`s evaluation | |
| * - measurement uncertainty – sampling and analysis – determined by extension coefficient k=2 (with probability of 95%). If sample is taken by the customer uncertainty of sampling is not available. | |
| - uncertainty given in % reflects the uncertainty from the result of measurement. | |
| ** - Acceptable to consumers and no abnormal change | |
| SL - analysis laboratory: NZ-Nové Zámky, TR-Turčianske Teplice, RK-Ružomberok, TV-Trebišov | |

Disclaimer: Laboratory is a disclaimer when the information is supplied by the customer (#) and can affect the validity of results. If the sample has been provided by the customer, the results refer to the sample as it was received. Gauges and measuring equipment used for testing were calibrated or attested in accordance with the valid metrological instructions. The above mentioned test results refer to the tested sample only! The result given in this Analytical report and marked as non accredited test shall not be a subject of accreditation. The result given in this Analytical report and marked as sub-delivery is the result of a Subcontractors gauging made under the terms and conditions of a contract concluded with him. This Analytical report shall not be reproduced except in full colour version, without written approval of the laboratory. SNAS is a Signatory to the Multilateral Agreement MRA ILAC.

Responsible for correctness:

Michaela Ruttkayová
Specialist worker

Worked out by: Zuzana Kubisová

Validity check of document



Test Certificate approved by

Michaela Ruttkayová
Specialist worker

